Chemical Change Record Sheet

WISBEY DENTAL
KNOWLEDGE | PASSION | SERVICE

NAME	DATE	CHEMICAL

For chemical supplies contact us on:
(02) 9875 5588

ASK ABOUT
VELOPEX VISCHECK STRIPS
SIMPLE CHEMICAL CONDITION TESTING

Wisbey Dental
265-271 Pennant Hills Road
Thornleigh NSW 2120

Tel: 02 9875 5588
Fax: 02 9875 5544

www.wisbeydental.com

IMPORTANT OPERATING INSTRUCTIONS

Remove lid, check levels are correct. Turn on water supply and mains switch. When amber ready light goes out processing can begin. Press process button to begin cycle and again before leaving processor.

WHEN CHANGING CHEMICALS

1. Drain tanks.
2. Remove modules and scrub thoroughly in hot water (NOT BOILING) including gears, with brush provided.
3. Fill developer and fixer tanks with cold water to overflow level and run system for 4 minutes. Then empty tanks completely.
4. Clean and wipe dry water tank then fill to operating level.
5. Fill tanks with fresh VELOPEX chemical to 1 cm below overflow and replace modules.
6. Top up chemicals to overflow if required.
7. Clean outside of machine.